

POLARIMETRY AND REFRACTOMETRY

Extremely Reliable Polarimeter for Practical Experience

THE STATE-OF-THE-ART POLARIMETER

Combining wide measuring range with simplicity

- » Entirely Made in Italy
- » Outstanding quality/price ratio
- » New design for a compact and space-saving solution

MONOCHROMATIC LED LIGHT SOURCE

Instead of traditional sodium bulb

- » Unparalleled illumination
- » 30x longer lifespan than conventional polarimeters
- » Plug & play, no heating time required



Polarimeter - POL-X

The POL-X polarimeter from Optika precisely determines the optical rotation of substances, quickly and easily. POL-X measures the optical rotation over the entire measuring range at the same high accuracy. It's suitable for many applications, from routine measurements to demanding experiments for innovative projects. With this new polarimeter, OPTIKA meets the requirements of the pharmaceutical, cosmetics, chemical and medical industries and creates an ideal tool for R&D applications.



Polarimetry is used in the Educational field (for physics & chemistry dept.) and for Quality & Process Control in Labs. It provides the concentration and purity of a substance before it is added to an expensive batch.



Optical rotation is an indispensable quality and identity assay for a wide range of critical industries. Food, drug, and flavors industries use polarimetry as a quality attribute both for raw ingredients and also finished products.

Polarimeter - POL-X

Technical Specifications

Measuring range of optical rotation: $\pm 180^\circ$

Resolution: 1°

Accuracy: 0.05°

Magnification factor of the glass: 4x

Light source: Monochromatic LED, 1.2 W, $\lambda_d = 590$ nm (equivalent to sodium lamp)

Length of test tubes: 100 mm and 200 mm

Weight: 1.7 Kg

Overall dimensions: 450x180x320 mm

Accessories:

POL-1.2 - 100 mm polarimeter tube

POL-1.3 - 200 mm polarimeter tube



Target Industries

Food

Concentration and purity of compounds (carbohydrates, lactose, raffinose, various starches, fructose, levulose, sucrose, natural monosaccharides, glucose, maltose, xylose) in sugar based foods, cereals and syrups.



Chemical

Identification and characterization of polymers (bio-, natural or synthetic), foods, cereals and syrups.



Pharmaceutical

Determination of product purity by measuring specific and optical rotation of amino acids, antibiotics, dextrose, steroids, amino sugars, cocaine, diuretics, tranquilizers, analgesics, codeine, serums, vitamins, etc.



Fragrance, Flavor & Essential Oils

Inspection of incoming raw materials, such as camphors, gums, orange oil, citric acid, lavender oil, spearmint oil, glyceric acid, lemon oil, etc.



Refractometers - HR & HRD Series




OPTIKA offers a wide range of hand refractometers to ensure incredibly rapid and convenient measurement of concentration in liquid and semi-solid samples, combining **accurate performance with excellent repeatability.**


The refractive index of a substance is related to its specific density; a refractometer is used to measure the purity or the concentration of a sample when mixed.

Refractometers are suitable for a wide range of applications, which includes the control of blend ratios in light industrial applications (such as glycols, battery acid, heat exchange fluids, coolants, quenchant and hydraulic oils) and represents the ideal solution for users working in the food industry (fruit, beverages, confectionery, jam, honey and other sugar based products).

- » **Easy operation**
- » **Built-in LED illumination**
- » **Brix or scale specific (including ATC)**
- » **Sturdy construction with rubber handgrip**



ATC - Automatic Temperature Compensation
No need to worry about temperature change during your measurement.





LED
Scale reading has never been so easy, also in case of low light conditions.



HR Series - Hand Refractometers



HR-110N - Hand Refractometer, 0-20% Brix, ATC, Built-in LED Illuminator

Measure Range	Min. Div	Accuracy
0-20% Brix	0.1 Brix	±0.10 Brix





HR-130N - Hand Refractometer, 0-32% Brix, ATC, Built-in LED Illuminator

Measure Range	Min. Div	Accuracy
0-32% Brix	0.2 Brix	±0.2 Brix





HR-140N - Hand Refractometer, 0-32% Brix, ATC, Triple Scale, Built-in LED Illuminator

Measure Range	Min. Div	Accuracy
0-32% Brix	0.2 Brix	±0.2 Brix
0-140 °Oe	1 °Oe	±1 °Oe
0-27 KMW(Babo)	0.2 KMW(Babo)	±0.2KMW(Babo)





HR-150N - Hand Refractometer, 0-80% Brix, Single Scale, Built-in LED Illuminator

Measure Range	Min. Div	Accuracy
0-80% Brix	0.5 Brix	±0.5 Brix





HR-160N - Hand Refractometer for Urine and Protein, ATC, Built-in LED Illuminator

Measure Range	Min. Div	Accuracy
1.000-1.050 sg	0.002 sg	± 0.002 sg
1.3300 RI - 1.3600 RI	0.0005 RI	± 0.0005 RI
0-12 g/dl	0.2 g/dl	± 0.2 g/dl





HR-170N - Hand Refractometer for Battery and Antifreeze, ATC, Built-in LED Illuminator

Measure Range	Min. Div	Accuracy
E: -60° F - 32° F	10° F	± 10°F
P: -50° F - 32° F	10° F	± 10°F
B: 1.15-1.30 sg	0.01 sg	± 0.1sg





HR-180N - Hand Refractometer for Wine and Grape, ATC, Built-in LED Illuminator

Measure Range	Min. Div	Accuracy
0-80% v/v	1,0% v/v	±1,0 v/v



HR-190N - Hand Refractometer, 0-28% Salinity, ATC, Built-in LED Illuminator


Measure Range	Min. Div	Accuracy
0-28% Salinity	0.2%	±0.2%



HRD Series - Digital Refractometers


HRD-300N - Digital Refractometer, Brix and Refractive Index

Measure Range	Min. Div	Accuracy
Brix 0-50%	0.10%	±0.2 %
Refractive index: 1.3330-1.4200nD	0.0001 nd	± 0.0003 nD




HRD-400N - Digital Refractometer, Brix, Salinity and Refractive Index

Measure Range	Min. Div	Accuracy
Brix 0-50%	± 0.1 %	± 0.2 %
Salinity 0.0-28.0%	± 0.1%	± 0.2 %
Refractive index: 1.3330-1.4200 nD	± 0.0001 nD	± 0.0003 nD



HRD-500N - Digital Refractometer for Urine, Serum Protein and Refractive Index

Measure Range	Min. Div	Accuracy
Urine SPG 1.000-1050	0.001	±0.002
SERUM P. 0-12 g/dl	0.1	±0.2
Refractive index: 1.3330-1.3900 nD	±0.0001 nD	±0.0003 nD



Refractometers - 2WAJ ABBE Bench Refractometer



Abbe bench-top refractometer measures refractive index ND and average color dispersion NF-NC of transparent and semi-transparent liquid or solid samples. It is a must-have equipment in factories, teaching institutes and science research centers related to oil, grease, pharmacy, painting, food, sugar-refining and geological industries, among the others. It is unique thanks to its ability in measuring also solid samples, such as film, glass, and other transparent materials.

- Main prism:** horizontal
- Secondary prism:** fitted on a hinge
- Refraction index scale:** Nd 1.300 – 1.700
- Precision:** Nd ± 0.0003
- Division:** Nd 0.0005
- Sugars scale:** 0-95% from Nd 1.300 – 1.530
- Precision:** 0-50% = 0.2%; 51-95% = 0.1%
- Division:** 0.25%
- Thermometer scale:** 0°C - 70°C, div. 0.1°C
- Weight:** 4 Kg
- Size:** 140x100x235 mm

v 6.5 - OPTIKA reserves the right to make corrections, modifications, enhancements, improvements and other changes to its products at any time without notice.

Headquarters and Manufacturing Facilities

OPTIKA® S.r.l. Via Rigla, 30 - 24010 Ponteranica (BG) - ITALY - Tel.: +39 035.571.392 - info@optikamicroscopes.com

Optika Sales branches

OPTIKA® Spain spain@optikamicroscopes.com
OPTIKA® China china@optikamicroscopes.com
OPTIKA® India india@optikamicroscopes.com

OPTIKA® USA usa@optikamicroscopes.com
OPTIKA® Central America america@optikamicroscopes.com